
Midmark M11 - v1.0.0
Total Cycles: 2698

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 198.4 F 3.8
4:00 208.4 F 8.4
6:00 222.7 F 10.8
8:00 225.1 F 8.3
10:00 237.4 F 12.2
12:00 240.2 F 11.1
14:00 237.6 F 9.3
16:00 251.8 F 16.2
18:00 263.5 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.5 F 29.1
2:00 271.6 F 28.3
3:00 272.5 F 29.1
4:00 271.5 F 28.3

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 231.8 F 0.0
10:00 226.5 F 0.0
15:00 227.7 F 0.0
20:00 228.5 F 0.0
25:00 227.7 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

12/20/2022 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____