
Midmark M11 - v1.0.5
Total Cycles: 1441

Name: RITTER2
Sterilizer ID: V2313228
08/31/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 178.4 F 3.0
4:00 193.7 F 7.8
6:00 212.8 F 13.3
8:00 225.8 F 8.4
10:00 236.5 F 11.8
12:00 240.7 F 12.1
14:00 240.6 F 11.2
16:00 253.9 F 18.0
18:00 265.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.9
2:00 271.8 F 29.1
3:00 272.1 F 29.8
4:00 272.5 F 29.8

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 229.4 F 0.0
10:00 232.9 F 0.0
15:00 233.7 F 0.0
20:00 222.6 F 0.0
25:00 220.5 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

08/31/2022 03:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____