

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1012

Name: RITTER 1  
Sterilizer ID: V2178198  
07/14/2021 01:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.1 F 0.0  
2:00 155.7 F 1.4  
4:00 160.3 F 4.1  
6:00 183.7 F 7.8  
8:00 216.2 F 12.3  
10:00 221.3 F 7.0  
12:00 230.9 F 12.1  
14:00 232.2 F 8.9  
16:00 238.1 F 10.2  
18:00 244.7 F 13.2  
20:00 256.9 F 19.6  
22:00 268.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.2 F 29.5  
2:00 273.2 F 29.8  
3:00 272.7 F 29.5  
4:00 272.3 F 29.3

Min 271.2 F 28.7  
Max 273.2 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 222.9 F 0.0  
10:00 235.5 F 0.0  
15:00 226.9 F 0.0  
20:00 223.4 F 0.0  
25:00 217.3 F 0.0  
30:00 198.3 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 23:27  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:01:54  
CYCLE COMPLETE

07/14/2021 02:19 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01