
Midmark M11 - v1.0.5
Total Cycles: 2779

Name: RITTER2
Sterilizer ID: V2313228
08/16/2023 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 152.2 F 0.0
4:00 173.7 F 0.0
6:00 198.3 F 0.4
8:00 217.1 F 4.4
10:00 220.3 F 8.5
12:00 232.9 F 12.9
14:00 237.1 F 9.8
16:00 238.1 F 9.7
18:00 244.2 F 12.3
20:00 246.0 F 13.4
22:00 257.7 F 19.7
24:00 268.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 29.1
2:00 271.5 F 28.9
3:00 272.0 F 29.1
4:00 271.4 F 28.7

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 225.4 F 0.0
10:00 206.0 F 0.0
15:00 187.8 F 0.0
20:00 184.0 F 0.0
25:00 181.3 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:15
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

08/16/2023 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____