
Midmark M11 - v1.0.0
Total Cycles: 3682

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2023 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 69.6 F 0.0
4:00 80.3 F 0.0
6:00 105.2 F 1.2
8:00 131.1 F 3.6
10:00 158.0 F 6.4
12:00 184.3 F 9.6
14:00 199.3 F 12.7
16:00 219.6 F 5.4
18:00 227.2 F 8.8
20:00 234.3 F 12.7
22:00 234.0 F 8.5
24:00 245.3 F 13.5
26:00 242.2 F 11.4
28:00 254.2 F 17.4
30:00 264.7 F 23.6
32:00 271.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.9 F 28.4
2:00 272.1 F 28.5
3:00 272.4 F 28.7
4:00 272.3 F 29.0

Min 271.6 F 28.4
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 215.4 F 0.0
10:00 217.6 F 0.0
15:00 217.5 F 0.0
20:00 206.8 F 0.0
25:00 213.2 F 0.0
30:00 209.2 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:10:42
CYCLE COMPLETE

08/25/2023 10:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____