
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000181
8:47 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:31
PRESSURIZING 4:31
STERILIZING 5:42
136.0 C 220kPa 5:42
136.4 C 226kPa 6:12
136.4 C 225kPa 6:42
136.4 C 226kPa 7:12
136.4 C 226kPa 7:42
136.4 C 226kPa 8:12
136.4 C 225kPa 8:42
136.3 C 224kPa 9:12
136.3 C 224kPa 9:12
VENTING 9:12
AIR DRYING 10:15
CYCLE COMPLETE 25:15

Digital Signature #
E01BB946605D3915
