
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006690
10:11 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:35
STERILIZING 3:40
136.4 C 222kPa 3:40
136.4 C 225kPa 4:10
136.4 C 223kPa 4:40
136.4 C 223kPa 5:10
136.0 C 221kPa 5:40
136.4 C 225kPa 6:10
136.4 C 222kPa 6:40
136.4 C 223kPa 7:10
136.4 C 223kPa 7:10
VENTING 7:10
AIR DRYING 8:07
CYCLE COMPLETE 8:13

Digital Signature #
8DD57A06DFC5E106
