
Midmark M11 - v1.0.5
Total Cycles: 2766

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	71.6 F	0.0
2:00	74.2 F	0.0
4:00	93.3 F	0.0
6:00	99.2 F	0.0
8:00	149.0 F	0.0
10:00	178.0 F	0.0
12:00	197.4 F	0.0
14:00	209.9 F	0.0
16:00	211.7 F	0.0
18:00	211.7 F	0.0
20:00	211.8 F	0.0
22:00	211.8 F	0.0
24:00	211.9 F	0.0
26:00	212.1 F	0.2
28:00	227.4 F	5.2

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 238F
CHAMBER PRESS 9.7PSI
** ITEMS NOT STERILE **

08/15/2023 08:31 AM