
Midmark M11 - v1.0.0
Total Cycles: 1334

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2021 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.2 F 0.0
2:00 93.9 F 0.0
4:00 110.6 F 0.0
6:00 141.9 F 0.1
8:00 178.4 F 0.2
10:00 197.3 F 2.8
12:00 214.3 F 5.7
14:00 223.4 F 9.0
16:00 236.2 F 12.7
18:00 234.0 F 8.2
20:00 244.3 F 13.0
22:00 241.0 F 10.6
24:00 236.5 F 8.6
26:00 250.6 F 15.1
28:00 261.6 F 21.3
30:00 271.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.8
1:00 272.4 F 28.6
2:00 271.8 F 27.9
3:00 272.4 F 28.6
4:00 272.0 F 28.1

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 221.9 F 0.0
10:00 193.9 F 0.0
15:00 236.7 F 0.0
20:00 219.6 F 0.0
25:00 203.6 F 0.0
30:00 234.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:39
CYCLE COMPLETE

11/03/2021 10:41 AM
ACCEPTED: X
REJECTED:
Operator: U01