
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001807
14:39 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:06
STERILIZING 4:10
136.4 C 222kPa 4:10
136.4 C 225kPa 4:40
136.4 C 225kPa 5:10
136.4 C 225kPa 5:40
136.2 C 222kPa 6:10
136.4 C 225kPa 6:40
136.3 C 224kPa 7:10
136.4 C 224kPa 7:40
136.4 C 224kPa 7:40
VENTING 7:40
AIR DRYING 8:36
CYCLE COMPLETE 9:04

Digital Signature #
EA6668E5A8AD8096
