
Midmark M11 - v1.0.0
Total Cycles: 3215

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2023 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 187.4 F 3.0
4:00 197.8 F 7.4
6:00 214.5 F 12.4
8:00 225.7 F 7.3
10:00 238.5 F 12.5
12:00 237.4 F 9.9
14:00 238.2 F 8.7
16:00 249.1 F 14.8
18:00 261.0 F 21.4
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.7 F 29.2
4:00 272.2 F 28.5

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 221.3 F 0.0
10:00 240.5 F 0.0
15:00 232.1 F 0.0
20:00 223.7 F 0.0
25:00 222.2 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

05/10/2023 04:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____