

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002758  
12:31 30/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:56  
STERILIZING 3:00  
136.4 C 222kPa 3:00  
136.3 C 222kPa 3:30  
136.4 C 223kPa 4:00  
136.4 C 224kPa 4:30  
136.1 C 220kPa 5:00  
136.4 C 224kPa 5:30  
136.1 C 221kPa 6:00  
136.4 C 223kPa 6:30  
136.4 C 223kPa 6:30  
VENTING 6:30  
AIR DRYING 7:23  
CYCLE COMPLETE 7:32

-----  
Digital Signature #  
33E0C54D48598772  
-----