
Midmark M11 - v1.0.0
Total Cycles: 1289

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.6 F 0.0
2:00 152.8 F 2.1
4:00 177.9 F 5.3
6:00 196.5 F 10.2
8:00 216.7 F 4.6
10:00 233.7 F 9.7
12:00 234.9 F 8.7
14:00 243.6 F 12.3
16:00 240.8 F 10.7
18:00 254.5 F 17.4
20:00 266.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 273.0 F 29.1
2:00 271.6 F 28.3
3:00 272.9 F 29.2
4:00 271.8 F 28.6

Min 271.4 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 197.9 F 0.0
10:00 194.6 F 0.0
15:00 196.4 F 0.0
20:00 195.7 F 0.0
25:00 193.9 F 0.0
30:00 191.8 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

10/22/2021 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____