
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003994
9:14 13/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:03
STERILIZING 3:06
136.4 C 224kPa 3:06
136.4 C 222kPa 3:36
136.3 C 224kPa 4:06
136.4 C 224kPa 4:36
136.1 C 219kPa 5:06
136.4 C 222kPa 5:36
136.0 C 220kPa 6:06
136.3 C 223kPa 6:36
136.3 C 223kPa 6:36
VENTING 6:37
AIR DRYING 8:11
CYCLE COMPLETE 10:20

Digital Signature #
718C9914590BEFD9
