

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5394

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2025 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.7 F 0.0  
2:00 135.5 F 1.6  
4:00 142.5 F 4.1  
6:00 166.1 F 7.4  
8:00 191.7 F 11.0  
10:00 212.1 F 4.3  
12:00 217.5 F 7.6  
14:00 230.4 F 12.0  
16:00 229.3 F 8.0  
18:00 240.4 F 13.2  
20:00 239.6 F 11.1  
22:00 250.9 F 16.8  
24:00 261.8 F 22.9  
26:00 271.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.3 F 30.1  
2:00 272.9 F 29.9  
3:00 272.1 F 29.5  
4:00 271.2 F 29.2

Min 271.2 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 203.6 F 0.0  
10:00 213.7 F 0.0  
15:00 197.4 F 0.0  
20:00 216.3 F 0.0  
25:00 221.9 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 26:42  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:05:26  
CYCLE COMPLETE

03/06/2025 12:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_