
Midmark M11 - v1.0.5
Total Cycles: 3591

Name: RITTER2
Sterilizer ID: V2313228
05/06/2024 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 74.4 F 0.0
4:00 122.9 F 0.3
6:00 106.3 F 1.8
8:00 138.0 F 4.1
10:00 161.6 F 6.4
12:00 189.0 F 8.7
14:00 197.1 F 12.2
16:00 214.0 F 4.8
18:00 219.8 F 8.2
20:00 233.5 F 11.8
22:00 229.0 F 7.3
24:00 239.1 F 12.1
26:00 237.2 F 9.8
28:00 248.8 F 15.3
30:00 258.9 F 21.0
32:00 268.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 271.6 F 29.6
2:00 272.0 F 29.6
3:00 272.4 F 29.8
4:00 272.1 F 30.0

Min 271.5 F 29.6
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 223.7 F 0.0
10:00 190.5 F 0.0
15:00 186.4 F 0.0
20:00 190.8 F 0.0
25:00 204.5 F 0.0
30:00 206.5 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:59
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:12:27
CYCLE COMPLETE

05/06/2024 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____