
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001402
9:56 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 5:18
STERILIZING 6:42
136.0 C 220kPa 6:42
136.2 C 221kPa 7:12
136.5 C 222kPa 7:43
136.4 C 222kPa 8:13
136.4 C 222kPa 8:43
136.4 C 223kPa 9:13
136.4 C 221kPa 9:43
136.4 C 223kPa 10:12
VENTING 10:12
AIR DRYING 11:12
CYCLE COMPLETE 11:16

Digital Signature #
E4356ADD141FCC92
