
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002518
14:47 19/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:47
STERILIZING 3:48
136.4 C 223kPa 3:48
136.4 C 221kPa 4:18
136.4 C 223kPa 4:48
136.3 C 223kPa 5:18
136.4 C 222kPa 5:48
136.1 C 222kPa 6:18
136.3 C 224kPa 6:48
136.2 C 224kPa 7:18
136.2 C 224kPa 7:18
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 8:15

Digital Signature #
9B6F2B9BC067F1E2
