
Midmark M11 - v1.0.0
Total Cycles: 1670

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 178.0 F 2.3
4:00 192.1 F 6.4
6:00 209.7 F 11.1
8:00 221.6 F 5.7
10:00 234.6 F 10.9
12:00 233.5 F 7.7
14:00 244.8 F 11.9
16:00 245.9 F 12.8
18:00 258.2 F 19.5
20:00 269.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 271.5 F 28.2
3:00 272.2 F 28.5
4:00 271.3 F 28.0

Min 271.3 F 27.9
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 214.8 F 0.0
10:00 229.7 F 0.0
15:00 225.2 F 0.0
20:00 230.0 F 0.0
25:00 226.4 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

02/24/2022 02:50 PM
ACCEPTED: X
REJECTED:
Operator: U02