
Midmark M11 - v1.0.5
Total Cycles: 893

Name: RITTER2
Sterilizer ID: V2313228
03/22/2022 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	72.0 F	0.0
2:00	85.9 F	0.0
4:00	102.0 F	0.0
6:00	109.5 F	0.0
8:00	156.5 F	0.0
10:00	189.5 F	0.0
12:00	211.1 F	0.0
14:00	213.0 F	0.1
16:00	213.1 F	0.2
18:00	213.1 F	0.2
20:00	213.2 F	0.2
22:00	213.2 F	0.2
24:00	213.1 F	0.2
26:00	213.1 F	0.2
28:00	213.1 F	0.2

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 213F
CHAMBER PRESS 0.1PSI
** ITEMS NOT STERILE **

03/22/2022 09:48 AM