

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1392

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2022 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.5 F 0.0  
2:00 156.8 F 0.0  
4:00 179.9 F 1.2  
6:00 199.9 F 5.6  
8:00 216.1 F 10.2  
10:00 224.0 F 7.4  
12:00 235.5 F 11.2  
14:00 235.2 F 9.0  
16:00 239.8 F 10.3  
18:00 248.0 F 14.5  
20:00 259.8 F 21.1  
22:00 270.6 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.4 F 29.0  
2:00 272.0 F 29.2  
3:00 272.2 F 28.9  
4:00 272.2 F 29.4

Min 271.6 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
5:00 201.1 F 0.0  
10:00 208.6 F 0.0  
15:00 209.2 F 0.0  
20:00 207.0 F 0.0  
25:00 204.4 F 0.0  
30:00 202.4 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:46  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:03  
CYCLE COMPLETE

08/15/2022 01:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_