
Midmark M11 - v1.0.5
Total Cycles: 3021

Name: RITTER2
Sterilizer ID: V2313228
10/17/2023 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 171.5 F 3.6
4:00 192.9 F 8.5
6:00 210.5 F 8.6
8:00 225.6 F 9.0
10:00 235.7 F 9.4
12:00 243.5 F 13.2
14:00 244.0 F 12.5
16:00 257.6 F 19.8
18:00 269.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.4 F 29.3
2:00 272.8 F 29.7
3:00 272.0 F 29.7
4:00 272.1 F 29.3

Min 271.6 F 29.1
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 239.9 F 0.0
10:00 225.7 F 0.0
15:00 225.5 F 0.0
20:00 230.1 F 0.0
25:00 234.9 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

10/17/2023 04:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____