
Midmark M11 - v1.0.0
Total Cycles: 3954

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 69.1 F 0.3
4:00 78.0 F 1.3
6:00 114.8 F 3.5
8:00 150.0 F 6.4
10:00 177.3 F 9.8
12:00 202.6 F 11.5
14:00 209.0 F 4.2
16:00 220.2 F 7.4
18:00 229.3 F 11.1
20:00 229.4 F 6.5
22:00 239.4 F 11.3
24:00 236.2 F 8.7
26:00 249.9 F 15.0
28:00 261.1 F 21.4
30:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.3
2:00 271.6 F 28.5
3:00 272.6 F 29.4
4:00 272.3 F 28.9

Min 271.4 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 221.6 F 0.0
10:00 202.3 F 0.0
15:00 193.2 F 0.0
20:00 192.6 F 0.0
25:00 195.8 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:41
CYCLE COMPLETE

10/30/2023 09:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____