
Midmark M11 - v1.0.5
Total Cycles: 3335

Name: RITTER2
Sterilizer ID: V2313228
01/30/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 73.9 F 0.0
4:00 113.0 F 0.0
6:00 104.4 F 1.0
8:00 139.6 F 3.8
10:00 166.1 F 6.8
12:00 182.9 F 10.6
14:00 210.6 F 6.5
16:00 217.2 F 6.2
18:00 226.1 F 9.9
20:00 235.8 F 10.0
22:00 237.2 F 10.6
24:00 235.6 F 8.4
26:00 248.2 F 14.7
28:00 259.3 F 20.9
30:00 269.3 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.5 F 29.5
2:00 271.9 F 29.4
3:00 271.6 F 28.8
4:00 272.3 F 29.3

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.0 F 0.0
5:00 233.3 F 0.0
10:00 218.2 F 0.0
15:00 181.8 F 0.0
20:00 177.5 F 0.0
25:00 174.9 F 0.0
30:00 178.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:55
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

01/30/2024 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____