

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4736

Name: RITTER 1  
Sterilizer ID: V2178198  
07/16/2024 03:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.6 F 0.0  
2:00 122.3 F 1.5  
4:00 140.1 F 4.3  
6:00 161.4 F 7.6  
8:00 187.2 F 11.4  
10:00 210.4 F 3.5  
12:00 216.6 F 7.9  
14:00 228.6 F 12.2  
16:00 230.4 F 8.2  
18:00 241.5 F 12.9  
20:00 241.6 F 11.8  
22:00 254.1 F 17.9  
24:00 264.9 F 24.3  
26:00 271.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.3 F 29.7  
2:00 272.4 F 29.4  
3:00 271.4 F 28.8  
4:00 271.3 F 29.2

Min 271.0 F 28.7  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 203.7 F 0.0  
10:00 200.8 F 0.0  
15:00 200.8 F 0.0  
20:00 202.4 F 0.0  
25:00 215.9 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:03  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:04:47  
CYCLE COMPLETE

07/16/2024 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_