
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000193
16:39 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:14
STERILIZING 2:54
136.4 C 222kPa 2:54
136.3 C 223kPa 3:24
136.4 C 222kPa 3:54
136.4 C 225kPa 4:24
136.5 C 224kPa 4:54
136.2 C 223kPa 5:24
136.0 C 221kPa 5:54
136.4 C 225kPa 6:24
136.4 C 225kPa 6:24
VENTING 6:24
AIR DRYING 7:12
CYCLE COMPLETE 7:14

Digital Signature #
1004A9AC2EE2F45B
