

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1009

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2022 07:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
2:00 180.3 F 0.3  
4:00 195.3 F 3.1  
6:00 217.4 F 8.3  
8:00 229.3 F 12.9  
10:00 236.7 F 10.0  
12:00 236.3 F 8.9  
14:00 242.7 F 11.4  
16:00 247.8 F 14.2  
18:00 260.2 F 21.1  
20:00 271.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.4 F 29.3  
2:00 272.1 F 28.7  
3:00 271.6 F 29.0  
4:00 272.7 F 29.1  
  
Min 271.2 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 232.2 F 0.0  
10:00 230.8 F 0.0  
15:00 231.2 F 0.0  
20:00 224.4 F 0.0  
25:00 229.0 F 0.0  
30:00 213.4 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:33  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:15  
CYCLE COMPLETE

04/20/2022 08:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_