
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005458
9:13 02/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:38
PRESSURIZING 1:47
STERILIZING 2:37
136.0 C 220kPa 2:37
136.4 C 225kPa 3:07
136.4 C 221kPa 3:37
135.8 C 220kPa 4:07
135.9 C 220kPa 4:37
136.4 C 222kPa 5:07
136.4 C 224kPa 5:37
136.1 C 220kPa 6:07
136.1 C 220kPa 6:07
VENTING 6:07
AIR DRYING 6:54
CYCLE COMPLETE 9:03

Digital Signature #
960EE3432897D54C
