
Midmark M11 - v1.0.0
Total Cycles: 3880

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2023 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 165.2 F 2.6
4:00 181.2 F 6.2
6:00 215.0 F 10.9
8:00 221.5 F 5.4
10:00 233.5 F 10.6
12:00 233.2 F 7.4
14:00 244.9 F 13.2
16:00 244.0 F 12.4
18:00 257.1 F 19.3
20:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.3 F 28.9
2:00 272.0 F 29.1
3:00 272.0 F 28.6
4:00 272.3 F 29.3

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 221.7 F 0.0
10:00 221.4 F 0.0
15:00 226.2 F 0.0
20:00 229.0 F 0.0
25:00 228.1 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

10/10/2023 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____