
Midmark M11 - v1.0.5
Total Cycles: 4287

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.5 F 0.0
2:00 149.4 F 1.8
4:00 159.9 F 5.1
6:00 189.0 F 9.3
8:00 223.0 F 8.7
10:00 224.5 F 7.4
12:00 238.1 F 11.9
14:00 234.3 F 8.7
16:00 242.1 F 11.2
18:00 245.3 F 13.2
20:00 256.9 F 19.5
22:00 267.6 F 26.1
24:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.6 F 29.4
2:00 273.1 F 29.7
3:00 271.9 F 29.8
4:00 271.9 F 29.3

Min 271.4 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 231.0 F 0.0
10:00 183.1 F 0.0
15:00 181.2 F 0.0
20:00 179.2 F 0.0
25:00 184.7 F 0.0
30:00 185.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

01/23/2025 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____