
Midmark M11 - v1.0.0
Total Cycles: 4073

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2023 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.3 F 0.0
2:00 155.5 F 1.6
4:00 171.8 F 4.7
6:00 178.2 F 8.3
8:00 199.2 F 12.2
10:00 211.5 F 4.5
12:00 222.2 F 8.4
14:00 231.2 F 12.7
16:00 233.0 F 8.7
18:00 242.0 F 12.0
20:00 243.7 F 12.3
22:00 255.2 F 18.2
24:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 272.1 F 28.8
3:00 271.2 F 28.3
4:00 271.8 F 29.0

Min 271.1 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 227.7 F 0.0
10:00 239.0 F 0.0
15:00 230.7 F 0.0
20:00 226.7 F 0.0
25:00 224.0 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

12/01/2023 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____