
Midmark M11 - v1.0.5
Total Cycles: 1224

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 07:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 75.2 F 0.0
4:00 86.1 F 0.0
6:00 106.4 F 0.0
8:00 151.2 F 0.3
10:00 176.9 F 2.7
12:00 193.7 F 6.1
14:00 210.8 F 9.7
16:00 221.8 F 13.4
18:00 227.0 F 7.8
20:00 238.0 F 12.1
22:00 235.7 F 9.4
24:00 238.6 F 9.8
26:00 246.9 F 14.1
28:00 258.1 F 20.4
30:00 268.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.9
2:00 272.1 F 29.2
3:00 271.5 F 29.1
4:00 273.0 F 29.9

Min 271.4 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 204.6 F 0.0
10:00 218.1 F 0.0
15:00 228.1 F 0.0
20:00 230.8 F 0.0
25:00 228.3 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

06/27/2022 08:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____