
Midmark M11 - v1.0.5
Total Cycles: 763

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.1 F 0.0
2:00 155.3 F 2.6
4:00 171.8 F 6.4
6:00 201.8 F 11.6
8:00 217.3 F 6.4
10:00 231.9 F 12.2
12:00 235.4 F 10.0
14:00 235.8 F 8.7
16:00 249.8 F 15.8
18:00 262.2 F 22.8
20:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 29.6
2:00 271.3 F 29.0
3:00 272.4 F 29.8
4:00 271.9 F 29.2

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 221.6 F 0.0
10:00 229.2 F 0.0
15:00 229.3 F 0.0
20:00 226.5 F 0.0
25:00 225.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

02/07/2022 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____