
Midmark M11 - v1.0.0
Total Cycles: 2728

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2023 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	123.9 F	0.0
2:00	127.8 F	0.0
4:00	144.1 F	1.0
6:00	173.7 F	3.8
8:00	196.1 F	8.3
10:00	212.8 F	13.1
12:00	223.0 F	6.7
14:00	234.8 F	11.3
16:00	233.6 F	8.0
18:00	244.2 F	12.2
20:00	244.9 F	12.6
22:00	257.8 F	19.4
24:00	269.4 F	26.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.0
1:00	272.5 F	28.5
2:00	271.6 F	28.2
3:00	272.4 F	28.5
4:00	271.6 F	28.2

Min 271.4 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	175.2 F	0.0
5:00	206.6 F	0.0
10:00	215.0 F	0.0
15:00	220.0 F	0.0
20:00	213.1 F	0.0
25:00	212.7 F	0.0
30:00	212.3 F	0.0

Drying Complete

FILLING: 0:48
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

01/03/2023 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____