

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1541

Name: RITTER 1  
Sterilizer ID: V2178198  
01/19/2022 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.4 F 0.0  
2:00 167.6 F 3.6  
4:00 181.1 F 8.0  
6:00 207.9 F 12.9  
8:00 219.8 F 7.1  
10:00 234.2 F 12.3  
12:00 235.4 F 9.5  
14:00 239.2 F 9.0  
16:00 248.2 F 14.6  
18:00 260.5 F 21.4  
20:00 271.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.3 F 29.1  
2:00 271.3 F 28.6  
3:00 272.4 F 29.2  
4:00 271.5 F 28.5  
  
Min 271.2 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 220.0 F 0.0  
10:00 231.5 F 0.0  
15:00 230.6 F 0.0  
20:00 228.9 F 0.0  
25:00 229.6 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:40  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:20  
CYCLE COMPLETE

01/19/2022 05:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_