
Midmark M11 - v1.0.0
Total Cycles: 1624

Name: RITTER 1
Sterilizer ID: V2178198
02/11/2022 03:33 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.8 F 0.0
2:00 78.9 F 0.0
4:00 90.9 F 0.5
6:00 122.8 F 1.9
8:00 152.8 F 3.7
10:00 179.4 F 6.9
12:00 198.3 F 10.9
14:00 213.8 F 4.2
16:00 225.3 F 7.4
18:00 236.8 F 11.7
20:00 233.8 F 8.1
22:00 244.1 F 12.1
24:00 244.4 F 12.3
26:00 257.0 F 19.0
28:00 268.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.6 F 28.1
2:00 271.7 F 28.1
3:00 272.0 F 28.3

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 229.3 F 0.0
5:00 214.3 F 0.0
10:00 204.7 F 0.0
15:00 203.0 F 0.0
20:00 201.2 F 0.0
25:00 199.6 F 0.0
30:00 200.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 29:07
STERILIZING: 3:00
VENTING: 2:49
DRYING: 30:00
TOTAL CYCLE: 01:06:03
CYCLE COMPLETE

02/11/2022 04:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____