
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000824
10:22 17/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:44
STERILIZING 3:48
136.0 C 227kPa 3:48
136.2 C 230kPa 4:18
136.4 C 231kPa 4:48
136.4 C 231kPa 5:18
136.4 C 229kPa 5:48
136.4 C 232kPa 6:18
136.4 C 232kPa 6:48
136.1 C 229kPa 7:18
136.1 C 229kPa 7:18
VENTING 7:18
AIR DRYING 8:11
CYCLE COMPLETE 23:11

Digital Signature #
EC60D0DBC226EBF3
