

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1815

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2022 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.6 F 0.0  
2:00 82.4 F 0.0  
4:00 115.9 F 0.0  
6:00 113.5 F 0.0  
8:00 143.1 F 0.0  
10:00 187.3 F 0.1  
12:00 197.1 F 3.8  
14:00 210.9 F 7.1  
16:00 225.3 F 10.7  
18:00 230.1 F 8.0  
20:00 237.3 F 10.7  
22:00 235.8 F 8.1  
24:00 243.5 F 11.9  
26:00 245.4 F 13.0  
28:00 257.2 F 19.4  
30:00 267.7 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.1  
2:00 272.0 F 29.2  
3:00 272.0 F 28.7  
4:00 272.5 F 29.5

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 212.3 F 0.0  
10:00 225.3 F 0.0  
15:00 229.5 F 0.0  
20:00 230.4 F 0.0  
25:00 221.0 F 0.0  
30:00 225.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:23  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:09:14  
CYCLE COMPLETE

12/07/2022 10:06 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_