

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1692

Name: RITTER 1  
Sterilizer ID: V2178198  
03/02/2022 11:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
2:00 186.6 F 3.0  
4:00 186.7 F 7.3  
6:00 212.9 F 12.6  
8:00 224.1 F 7.4  
10:00 237.3 F 12.8  
12:00 236.5 F 9.9  
14:00 236.1 F 8.5  
16:00 248.6 F 14.7  
18:00 260.7 F 21.3  
20:00 271.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.3  
2:00 272.1 F 28.8  
3:00 271.3 F 28.4  
4:00 272.2 F 29.2

Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 226.9 F 0.0  
10:00 232.5 F 0.0  
15:00 235.1 F 0.0  
20:00 225.1 F 0.0  
25:00 224.8 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 20:40  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

03/02/2022 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_