
Midmark M11 - v1.0.0
Total Cycles: 2643

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2022 08:17 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.4 F 0.0
2:00 187.1 F 3.6
4:00 203.0 F 8.2
6:00 218.5 F 10.7
8:00 230.6 F 9.4
10:00 235.0 F 9.0
12:00 244.4 F 13.0
14:00 243.6 F 12.0
16:00 257.1 F 19.1
18:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.2 F 28.8
2:00 271.2 F 28.5
3:00 272.2 F 28.8

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 229.0 F 0.0
10:00 228.6 F 0.0
15:00 230.8 F 0.0
20:00 228.3 F 0.0
25:00 230.8 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 19:03
STERILIZING: 3:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:55
CYCLE COMPLETE

12/05/2022 09:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____