
Midmark M11 - v1.0.0
Total Cycles: 4417

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2024 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.8 F 0.0
2:00 92.7 F 0.7
4:00 114.6 F 2.4
6:00 148.7 F 5.5
8:00 176.2 F 9.3
10:00 222.0 F 11.1
12:00 222.1 F 5.5
14:00 231.4 F 9.2
16:00 239.0 F 13.3
18:00 235.8 F 9.1
20:00 240.2 F 10.8
22:00 243.4 F 12.0
24:00 255.2 F 18.0
26:00 265.8 F 24.3
28:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.7 F 29.0
2:00 271.9 F 29.0
3:00 272.4 F 28.8
4:00 271.8 F 29.0

Min 271.3 F 28.4
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 235.6 F 0.0
10:00 190.7 F 0.0
15:00 200.1 F 0.0
20:00 205.7 F 0.0
25:00 217.0 F 0.0
30:00 208.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:29
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

03/28/2024 02:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____