

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1247

Name: RITTER 1  
Sterilizer ID: V2178198  
10/08/2021 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.1 F 0.0  
2:00 141.5 F 0.7  
4:00 162.6 F 3.5  
6:00 184.4 F 6.9  
8:00 204.4 F 11.5  
10:00 219.5 F 5.2  
12:00 231.4 F 10.0  
14:00 233.7 F 8.2  
16:00 242.1 F 11.8  
18:00 238.4 F 9.5  
20:00 251.4 F 15.9  
22:00 262.3 F 22.4  
24:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 29.1  
2:00 272.0 F 28.4  
3:00 272.1 F 28.8  
4:00 272.4 F 29.0

Min 271.2 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
5:00 226.5 F 0.0  
10:00 224.2 F 0.0  
15:00 211.9 F 0.0  
20:00 200.3 F 0.0  
25:00 205.6 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:24  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

10/08/2021 09:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_