

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3102

Name: RITTER 1  
Sterilizer ID: V2178198  
04/12/2023 03:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.7 F 0.0  
2:00 118.1 F 0.0  
4:00 138.3 F 0.5  
6:00 176.6 F 3.1  
8:00 197.3 F 7.1  
10:00 210.7 F 11.5  
12:00 220.3 F 5.2  
14:00 230.5 F 9.8  
16:00 237.0 F 9.9  
18:00 241.4 F 11.7  
20:00 238.6 F 9.8  
22:00 251.8 F 16.2  
24:00 263.0 F 22.6  
26:00 272.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 28.9  
2:00 271.4 F 28.1  
3:00 272.4 F 29.1  
4:00 272.0 F 28.6

Min 271.2 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 215.8 F 0.0  
10:00 222.0 F 0.0  
15:00 218.2 F 0.0  
20:00 219.6 F 0.0  
25:00 223.4 F 0.0  
30:00 224.2 F 0.0

Drying Complete

FILLING: 0:58  
HEATING: 26:18  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:04:39  
CYCLE COMPLETE

04/12/2023 04:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_