
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003172
14:40 16/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:02
STERILIZING 4:04
136.4 C 220kPa 4:04
136.5 C 225kPa 4:34
136.3 C 224kPa 5:04
136.3 C 225kPa 5:34
136.3 C 223kPa 6:04
136.4 C 224kPa 6:34
136.4 C 223kPa 7:04
136.4 C 222kPa 7:34
136.4 C 222kPa 7:34
VENTING 7:34
AIR DRYING 8:31
CYCLE COMPLETE 8:34

Digital Signature #
62B435416BD5D3ED
