
Midmark M11 - v1.0.0
Total Cycles: 4010

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2023 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.8 F 0.0
2:00 167.4 F 1.8
4:00 174.5 F 5.0
6:00 195.1 F 9.8
8:00 216.4 F 6.3
10:00 223.7 F 8.4
12:00 235.0 F 13.1
14:00 236.2 F 9.8
16:00 237.5 F 9.2
18:00 247.0 F 13.8
20:00 258.5 F 19.9
22:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.4
3:00 271.7 F 28.7
4:00 272.3 F 29.0

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 214.2 F 0.0
10:00 224.4 F 0.0
15:00 221.5 F 0.0
20:00 212.7 F 0.0
25:00 215.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

11/13/2023 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____