
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002244
11:02 12/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:55
STERILIZING 3:59
136.0 C 221kPa 3:59
136.3 C 224kPa 4:29
136.5 C 226kPa 4:59
136.2 C 221kPa 5:29
136.2 C 222kPa 5:59
136.3 C 224kPa 6:29
136.4 C 225kPa 6:59
136.5 C 225kPa 7:29
136.5 C 225kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 8:24

Digital Signature #
738BDD3303C0E34D
