

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4352

Name: RITTER 1  
Sterilizer ID: V2178198  
03/01/2024 11:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.9 F 0.0  
2:00 135.8 F 1.7  
4:00 146.3 F 4.4  
6:00 205.7 F 7.7  
8:00 206.2 F 12.1  
10:00 218.7 F 4.2  
12:00 225.2 F 8.2  
14:00 235.2 F 12.5  
16:00 235.1 F 8.5  
18:00 246.0 F 13.2  
20:00 243.4 F 11.6  
22:00 254.6 F 17.4  
24:00 264.6 F 23.3  
26:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 27.9  
1:00 272.4 F 28.9  
2:00 271.3 F 28.4  
3:00 272.0 F 28.4  
4:00 272.6 F 28.8

Min 270.6 F 27.8  
Max 273.0 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
5:00 189.3 F 0.0  
10:00 237.9 F 0.0  
15:00 230.8 F 0.0  
20:00 192.0 F 0.0  
25:00 187.6 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 26:03  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:43  
CYCLE COMPLETE

03/01/2024 01:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_