
Midmark M11 - v1.0.0
Total Cycles: 3089

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 175.4 F 3.2
4:00 183.9 F 7.1
6:00 209.6 F 12.1
8:00 221.0 F 6.6
10:00 234.6 F 12.0
12:00 234.8 F 9.0
14:00 238.2 F 9.5
16:00 247.1 F 14.0
18:00 259.5 F 20.7
20:00 270.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.3 F 28.4
3:00 272.5 F 29.2
4:00 271.6 F 28.5

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 217.8 F 0.0
10:00 228.0 F 0.0
15:00 230.1 F 0.0
20:00 226.7 F 0.0
25:00 224.1 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

04/11/2023 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____