
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002455
10:46 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:58
STERILIZING 4:06
136.4 C 224kPa 4:06
136.4 C 225kPa 4:36
136.4 C 225kPa 5:06
136.3 C 224kPa 5:36
136.4 C 225kPa 6:06
136.3 C 224kPa 6:36
136.3 C 224kPa 7:06
136.4 C 221kPa 7:36
136.4 C 221kPa 7:36
VENTING 7:37
AIR DRYING 8:31
CYCLE COMPLETE 25:05

Digital Signature #
0535167FBEEC5CDA
