
Midmark M11 - v1.0.5
Total Cycles: 1790

Name: RITTER2
Sterilizer ID: V2313228
11/30/2022 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 171.2 F 0.1
4:00 180.0 F 0.6
6:00 209.4 F 3.9
8:00 227.6 F 9.3
10:00 234.2 F 9.1
12:00 240.2 F 12.5
14:00 240.5 F 11.5
16:00 240.9 F 11.3
18:00 255.0 F 18.5
20:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.6
2:00 271.6 F 28.9
3:00 272.7 F 29.8
4:00 272.1 F 29.1

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 215.1 F 0.0
10:00 228.3 F 0.0
15:00 225.7 F 0.0
20:00 225.4 F 0.0
25:00 222.6 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

11/30/2022 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____